

# RIGI-L2 Camera



## Uncooled real-time THz imager



- Beam profiling
- Non-destructive testing
- Imaging systems
- Quality control
- Security and surveillance
- Optical alignment

## Image using QCL THz source

Frequency: 4.6 THz
Power: 0.6 μ W
Integration time: 50 μs

• **Acquisition:** Single shot

	ltem	General Specifications	
	Detector	Uncooled FPA THz micro-bolometer array	
	Pitch	15 µm	
	Array size	1280 x 1024 pixels	
	Act <mark>ive ar</mark> ea	19 x 15.4 mm	
	Operating fr	equency 0.1 - 20 THz	
	Pixel NEP	< 1.5 pW/√Hz at 4.6 THz	
	Operating to	emperature -40 °C to 65 °C	
	Software	Included and adaptable upon request	

### Compatible with most terahertz sources

### 1. Broadband terahertz sources

- Broadband plasma based source
- Organic crystal high power spectrometer
- Lithium Niobate crystal high power spectrometer
- ZnTe crystal based linear spectrometer
- Photoconductive switch oscillator based source
- Commercial time domain spectrometer

### 2. Narrowband terahertz sources

- OCL based source
- Solid state sources
- Crystal based DFG source
- CO2-based DFG source

Add: Floor 3, F Building, No. 86 Caobao road, Shanghai 200235 P.R. China

Tel: +86-21-51083793 Fax:+86-21-34241962

E-Mail: info@auniontech.com Website: www.auniontech.com